


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS CALIBRATION 9617 Accredited to ISO/IEC 17025:2005	MWS Limited	
	Issue No: 001 Issue date: 20 April 2017	
	Unit 6, Meer End Birstall Leicester LE4 3EH United Kingdom	Contact: Mr Paul Cooke Tel: +44 (0) 845 260 2602 Fax: +44 (0) 116 227 1316 E-Mail: paul.cooke@mws.ltd.uk Website: www.mws.ltd.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Unit 6, Meer End Birstall Leicester LE4 3EH United Kingdom	Local contact Mr Paul Cooke	Calibration of Non Automatic Weighing Machine P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At Customers Premises The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Calibration Engineer Calibration of Non Automatic Weighing Machine	S



9617
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Tres, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

MWS Limited
Issue No: 001 Issue date: 20 April 2017

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (<i>k</i> =2)	Remarks	Location Code
WEIGHING INSTRUMENTS Electronic digital one pan non- automatic weighing	100 g 200 g 500 g 1 kg 2 kg 5 kg 6 kg 10 kg 20 kg 50 kg 100 kg 200 kg 500 kg	0.77 mg 1.5 mg 3.9 mg 7.7 mg 15 mg 39 mg 46 mg 0.76 g 1.5 g 3.8 g 7.7 g 16 g 39 g	1. Weights are available in OIML Class: F1 from 1 g mg to 2 kg, Max. grouped load 6.1 kg M1 from 100 g to 20 kg. Max grouped load 500 kg. 2. Other loads within the overall listed range may also be used.	S
END				